



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/642,365

08/14/2003

Michael S. H. Chu

7062104001

1647

23517 7590 08/16/2011
BINGHAM MCCUTCHEN LLP
2020 K Street, N.W.
Intellectual Property Department
WASHINGTON, DC 20006

EXAMINER

RYCKMAN, MELISSA K

ART UNIT

PAPER NUMBER

3773

MAIL DATE

DELIVERY MODE

08/16/2011

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No.	Applicant(s)	
	10/642,365	CHU ET AL.	
	Examiner	Art Unit	
	MELISSA RYCKMAN	3773	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 8/3/11.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-9, 11-28 and 30-52 is/are pending in the application.
- 4a) Of the above claim(s) 1-8, 12 and 14-43 is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 9, 11, 13 and 44-52 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
 Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
 Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

This office action is in response to arguments filed 8/3/11.

Priority

As previously stated, the claims are given priority to 8/14/02, as the earlier dates of the CIP applications do not include the first and second legs as stated in the claims.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 9, 11, 13 and 44-52 are rejected under 35 U.S.C. 103(a) as being unpatentable over Rioux (U.S. Patent No. 7,527,633) and further in view of and Skiba et al. (U.S. Patent No. 6,723,107) and Anderson et al. (U.S. Pub. No. 2002/0099259).

Rioux teaches an implant placement device comprising the following:

- A slotted connector (26) formed in a distal end (20) of a shaft (24) of a delivery device (10, Fig. 1)
- A device for holding a suture material comprising a slotted connector (Fig. 2A), wherein the slotted connector includes first and second legs, the first leg extending radially into the shaft (near number 30, Fig. 2A) and the second leg (near number 26 Fig. 2A) extending axially in a distal direction


Art Unit: 3773

from the first leg along the shaft and being longer than the first leg and includes a narrowing (distal end of 26 is narrow) in order to permit easy insertion of the suture into the opening and to wedge the suture into tight frictional engagement with the device (Fig. 2A).

- Wherein the second leg is longer than the first leg (Rioux, Fig. 2A)
- The second leg extends distally at a 90 degree angle to the first leg (Rioux, Fig. 2A)
- The second leg of the slotted connector extends axially into the tapered section of the shaft (Rioux, Fig. 2A)
- The width of the narrowing is less than that of the loop material (the width at the most distal end of the second leg of Rioux is less than the width of the loop portion, Fig. 2A)
- The first leg of the slotted connector is proximal of the tapered section of the shaft (Rioux, Fig. 2A)
- The distal tip is solid (Rioux, Fig. 2A)
- The first and second legs join at an inside corner and an outside corner (Rioux, Fig. 2A), the outside corner being located proximal to the inside corner along the shaft
- The distal end of the shaft includes a tapered section that is substantially straight and conical (col. 10, ll. 62,63)
- The tapered section of the shaft is substantially straight (Fig. 1)
- The distal tip is blunt (col. 10, ll. 62,63)

Rioux teaches the claimed invention including the medical implant being a sling (col. 3, ll. 12), but does not specify the sling having a loop connector at the end. However, Anderson teaches, in the same field of endeavor, a sling having a loop connector at the end (Fig. 4, 48 and 50) for the purposes of delivering the sling. It would have been obvious to one of ordinary skill in the art at the time the invention was made to use the sling of Anderson with the device of Rioux as this is appropriate for use in the body, the loops aid in the completion of the procedure.

The combination of Anderson and Rioux is silent regarding the outside corner is curved and the inside corner is not curved, however Skiba teaches, in the same field of endeavor, using curved and not curved corners in the embodiment shown in Fig. 16 (curved/not curved, see Figure below that has been added for clarification), Skiba specifies a curved configuration aids in trapping a suture (col. 4, ll. 34). It would have been obvious to one of ordinary skill in the art at the time the invention was made to combine the embodiments of Skiba with the device of Anderson and Rioux to have one corner curved and the other not curved as this aids in placement of the loop and in using the device.



Response to Arguments

The applicant argues Skiba does not teach the specified corners of the connector. The examiner included Fig. 16 of Skiba (as shown above) to clarify how Skiba teaches the limitations of the outside corner is curved and the inside corner is not curved.

Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any

Art Unit: 3773

extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Melissa Ryckman whose telephone number is (571)272-9969 and email is Melissa.Ryckman@uspto.gov. The examiner can normally be reached on Monday-Friday 9:00-5:00.

If attempts to reach the examiner by telephone are unsuccessful, please contact the examiner's supervisor, Corrine McDermott, at (571) 272-4754. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

If there are any inquiries that are not being addressed by first contacting the Examiner or the Supervisor, you may send an email inquiry to

TC3700_Workgroup_D_Inquiries@uspto.gov.

Art Unit: 3773

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

MKR
/Melissa Ryckman/
Examiner, Art Unit 3773

/S. Thomas Hughes/
Supervisory Patent Examiner, Art Unit 3731